

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002545  
10:43 28/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:48  
STERILIZING 3:53  
136.5 C 224kPa 3:53  
136.4 C 226kPa 4:23  
136.1 C 223kPa 4:53  
136.4 C 225kPa 5:23  
136.4 C 225kPa 5:53  
136.1 C 223kPa 6:23  
136.1 C 222kPa 6:53  
136.2 C 224kPa 7:23  
136.2 C 224kPa 7:23  
VENTING 7:23  
AIR DRYING 8:13  
CYCLE COMPLETE 13:00

-----  
Digital Signature #  
6F2930E9ED7C55E5  
-----