
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001994
13:30 09/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:13
STERILIZING 3:11
136.4 C 222kPa 3:11
136.1 C 223kPa 3:41
136.2 C 224kPa 4:11
136.3 C 225kPa 4:41
136.3 C 224kPa 5:11
136.2 C 224kPa 5:41
136.4 C 224kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:32
CYCLE COMPLETE 37:33

Digital Signature #
971B6F9D965CC989
