
Midmark M11 - v1.0.5
Total Cycles: 4254

Name: RITTER2
Sterilizer ID: V2313228
01/10/2025 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 77.1 F 0.0
4:00 113.7 F 0.0
6:00 113.0 F 0.1
8:00 157.0 F 0.2
10:00 185.6 F 0.8
12:00 195.6 F 3.4
14:00 214.2 F 6.6
16:00 227.0 F 10.0
18:00 234.4 F 10.5
20:00 234.2 F 9.6
22:00 240.1 F 10.2
24:00 244.2 F 12.9
26:00 243.0 F 11.9
28:00 255.6 F 18.5
30:00 266.8 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.0 F 28.6
2:00 272.4 F 28.9
3:00 272.7 F 29.2
4:00 272.5 F 29.4

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 226.1 F 0.0
10:00 232.8 F 0.0
15:00 228.7 F 0.0
20:00 235.2 F 0.0
25:00 229.4 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

01/10/2025 10:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____