
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000400
15:39 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:04
STERILIZING 4:10
136.0 C 220kPa 4:10
136.1 C 231kPa 4:40
136.4 C 232kPa 5:10
136.3 C 232kPa 5:40
136.4 C 233kPa 6:10
136.3 C 233kPa 6:40
136.3 C 232kPa 7:10
136.4 C 233kPa 7:40
136.4 C 233kPa 7:40
VENTING 7:40
AIR DRYING 8:34
CYCLE COMPLETE 8:35

Digital Signature #
B03EBE422EB9CE33
