

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4347

Name: RITTER 1  
Sterilizer ID: V2178198  
02/29/2024 02:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.3 F 0.0  
2:00 113.2 F 1.2  
4:00 136.8 F 3.6  
6:00 170.5 F 7.0  
8:00 202.9 F 11.8  
10:00 216.0 F 3.9  
12:00 224.5 F 7.9  
14:00 235.3 F 12.2  
16:00 232.6 F 8.1  
18:00 244.4 F 13.5  
20:00 242.5 F 11.5  
22:00 254.4 F 17.6  
24:00 265.1 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.2  
1:00 272.5 F 29.0  
2:00 271.4 F 28.1  
3:00 272.6 F 28.9  
4:00 271.5 F 28.2

Min 271.3 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 234.5 F 0.0  
10:00 226.9 F 0.0  
15:00 227.5 F 0.0  
20:00 225.3 F 0.0  
25:00 220.0 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:43  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:04:04  
CYCLE COMPLETE

02/29/2024 03:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_