

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2577

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2023 07:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 77.7 F 0.0  
4:00 84.9 F 0.0  
6:00 115.9 F 0.0  
8:00 159.3 F 0.6  
10:00 185.2 F 3.9  
12:00 198.6 F 7.8  
14:00 225.6 F 11.9  
16:00 223.0 F 5.6  
18:00 232.7 F 9.7  
20:00 238.1 F 10.7  
22:00 241.0 F 11.6  
24:00 238.5 F 10.0  
26:00 251.5 F 16.4  
28:00 262.8 F 22.9  
30:00 272.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.7  
1:00 272.3 F 29.4  
2:00 272.1 F 28.8  
3:00 272.1 F 29.3  
4:00 272.4 F 29.1  
  
Min 271.5 F 28.5  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
5:00 222.0 F 0.0  
10:00 234.5 F 0.0  
15:00 226.1 F 0.0  
20:00 224.0 F 0.0  
25:00 217.7 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:12  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:08:10  
CYCLE COMPLETE

06/28/2023 08:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_