
Midmark M11 - v1.0.5
Total Cycles: 2226

Name: RITTER2
Sterilizer ID: V2313228
04/10/2023 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.1 F 0.0
2:00 159.8 F 2.9
4:00 175.1 F 6.6
6:00 199.0 F 11.9
8:00 219.6 F 6.8
10:00 234.2 F 12.3
12:00 235.1 F 9.6
14:00 236.4 F 9.0
16:00 248.3 F 14.9
18:00 260.7 F 21.8
20:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 29.2
2:00 272.3 F 29.8
3:00 271.1 F 28.9
4:00 273.1 F 30.2

Min 270.9 F 28.9
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 234.0 F 0.0
10:00 226.0 F 0.0
15:00 222.1 F 0.0
20:00 217.6 F 0.0
25:00 219.7 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

04/10/2023 05:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____