
Midmark M11 - v1.0.0
Total Cycles: 777

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2021 03:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.4 F 0.0
2:00 131.7 F 1.1
4:00 154.9 F 3.7
6:00 186.1 F 7.8
8:00 206.6 F 13.4
10:00 224.7 F 7.1
12:00 236.9 F 12.2
14:00 235.2 F 8.9
16:00 240.1 F 9.8
18:00 246.1 F 13.2
20:00 258.7 F 19.9
22:00 269.6 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.3 F 28.7
2:00 272.6 F 29.1
3:00 272.4 F 29.3
4:00 271.8 F 28.6

Min 271.6 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 235.6 F 0.0
10:00 226.6 F 0.0
15:00 215.8 F 0.0
20:00 214.3 F 0.0
25:00 231.2 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:41
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

04/07/2021 04:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____