

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3448

Name: RITTER 1  
Sterilizer ID: V2178198  
06/26/2023 03:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
2:00 193.0 F 3.1  
4:00 202.4 F 7.2  
6:00 216.6 F 12.5  
8:00 227.0 F 7.4  
10:00 239.2 F 12.5  
12:00 238.8 F 10.3  
14:00 236.8 F 8.5  
16:00 250.5 F 15.5  
18:00 262.2 F 22.2  
20:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.6 F 29.0  
2:00 271.5 F 28.3  
3:00 272.2 F 29.1  
4:00 272.2 F 28.9

Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
5:00 231.5 F 0.0  
10:00 199.9 F 0.0  
15:00 194.1 F 0.0  
20:00 190.0 F 0.0  
25:00 187.2 F 0.0  
30:00 185.2 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 00:59:52  
CYCLE COMPLETE

06/26/2023 04:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_