

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000108  
8:16 03/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:49  
PRESSURIZING 4:50  
STERILIZING 6:19  
136.4 C 226kPa 6:19  
136.2 C 225kPa 6:49  
136.3 C 225kPa 7:19  
136.3 C 225kPa 7:49  
136.4 C 224kPa 8:19  
136.2 C 223kPa 8:49  
136.4 C 226kPa 9:19  
136.2 C 223kPa 9:49  
136.2 C 223kPa 9:49  
VENTING 9:49  
AIR DRYING 10:55  
CYCLE COMPLETE 25:56  
-----

Digital Signature #  
C53F36F46D9FFA3B  
-----