
Midmark M11 - v1.0.5
Total Cycles: 2307

Name: RITTER2
Sterilizer ID: V2313228
04/27/2023 12:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.8 F 0.0
2:00 164.7 F 3.1
4:00 185.4 F 7.4
6:00 204.7 F 12.9
8:00 221.2 F 8.0
10:00 234.8 F 12.7
12:00 238.4 F 11.2
14:00 238.1 F 10.0
16:00 251.8 F 16.8
18:00 263.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 271.9 F 29.2
2:00 272.4 F 29.6
3:00 271.6 F 29.5
4:00 271.7 F 29.2

Min 271.2 F 29.0
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 206.8 F 0.0
10:00 218.3 F 0.0
15:00 224.8 F 0.0
20:00 227.2 F 0.0
25:00 227.4 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

04/27/2023 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____