

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2103

Name: RITTER2  
Sterilizer ID: V2313228  
02/28/2023 04:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.8 F 0.0  
2:00 162.4 F 3.5  
4:00 186.9 F 8.0  
6:00 207.3 F 13.4  
8:00 229.1 F 7.9  
10:00 235.2 F 12.4  
12:00 239.0 F 11.4  
14:00 238.5 F 10.4  
16:00 252.6 F 17.5  
18:00 265.2 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.5 F 29.8  
2:00 271.4 F 29.0  
3:00 272.3 F 30.0  
4:00 272.3 F 29.5  
  
Min 271.2 F 29.0  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
5:00 232.8 F 0.0  
10:00 228.6 F 0.0  
15:00 228.4 F 0.0  
20:00 231.2 F 0.0  
25:00 231.9 F 0.0  
30:00 229.0 F 0.0  
  
Drying Complete

FILLING: 1:14  
HEATING: 19:43  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:12  
CYCLE COMPLETE

02/28/2023 04:59 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_