

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4154

Name: RITTER2  
Sterilizer ID: V2313228  
11/22/2024 01:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.5 F 0.0  
2:00 167.1 F 2.3  
4:00 185.1 F 5.8  
6:00 212.0 F 9.8  
8:00 226.8 F 8.7  
10:00 228.1 F 8.1  
12:00 234.1 F 12.6  
14:00 235.4 F 9.8  
16:00 238.7 F 9.8  
18:00 246.6 F 14.2  
20:00 257.6 F 20.2  
22:00 267.6 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.3  
1:00 272.6 F 30.1  
2:00 272.4 F 29.9  
3:00 272.7 F 30.2  
4:00 272.4 F 30.4

Min 270.9 F 29.2  
Max 273.3 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 213.3 F 0.0  
10:00 218.6 F 0.0  
15:00 217.4 F 0.0  
20:00 219.4 F 0.0  
25:00 217.4 F 0.0  
30:00 213.7 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 23:30  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:02:53  
CYCLE COMPLETE

11/22/2024 02:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_