

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 001602  
14:15 11/19/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:44  
PRESSURIZING 5:27  
STERILIZING 6:55  
136.0 C 220kPa 6:55  
136.2 C 217kPa 7:25  
136.3 C 221kPa 7:55  
135.9 C 221kPa 8:25  
136.4 C 223kPa 8:55  
135.9 C 217kPa 9:25  
136.4 C 223kPa 9:55  
136.3 C 221kPa 10:25  
VENTING 10:25  
AIR DRYING 11:29  
CYCLE COMPLETE 41:29  
-----

Digital Signature #  
51BDBF6FA48B96C4  
-----