

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2346

Name: RITTER 1  
Sterilizer ID: V2178198  
09/20/2022 03:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.0 F 0.0  
2:00 173.4 F 3.5  
4:00 183.8 F 7.9  
6:00 213.2 F 10.9  
8:00 227.5 F 8.5  
10:00 237.1 F 10.9  
12:00 240.7 F 11.5  
14:00 239.1 F 10.1  
16:00 253.5 F 17.3  
18:00 265.8 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.6 F 29.2  
2:00 271.6 F 28.5  
3:00 271.9 F 29.0  
4:00 272.4 F 29.0  
  
Min 271.1 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 235.7 F 0.0  
10:00 224.2 F 0.0  
15:00 227.9 F 0.0  
20:00 229.4 F 0.0  
25:00 228.9 F 0.0  
30:00 230.6 F 0.0  
  
Drying Complete

FILLING: 1:58  
HEATING: 19:38  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

09/20/2022 04:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_