
Midmark M11 - v1.0.0
Total Cycles: 1649

Name: RITTER 1
Sterilizer ID: V2178198
02/17/2022 12:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.8 F 0.0
2:00 183.0 F 3.3
4:00 202.8 F 7.7
6:00 222.7 F 13.1
8:00 226.3 F 7.7
10:00 240.0 F 13.4
12:00 240.6 F 11.1
14:00 238.2 F 9.6
16:00 252.1 F 16.3
18:00 263.8 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 28.8
2:00 271.5 F 28.6
3:00 272.2 F 28.7
4:00 271.2 F 28.5

Min 271.0 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 255.6 F 0.0
10:00 227.1 F 0.0
15:00 223.2 F 0.0
20:00 231.0 F 0.0
25:00 230.2 F 0.0
30:00 227.5 F 0.0

Drying Complete

FILLING: 1:45
HEATING: 19:60
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

02/17/2022 01:49 PM
ACCEPTED: X
REJECTED:
Operator: U02