
Midmark M11 - v1.0.5
Total Cycles: 184

Name: RITTER2
Sterilizer ID: V2313228
07/08/2021 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.9 F 0.0
2:00 169.8 F 3.8
4:00 194.7 F 9.0
6:00 213.3 F 6.3
8:00 228.1 F 10.7
10:00 234.6 F 8.8
12:00 238.1 F 9.7
14:00 249.7 F 15.5
16:00 262.5 F 23.1
18:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.3 F 29.3
2:00 272.3 F 29.8
3:00 271.7 F 29.0
4:00 272.4 F 29.8

Min 271.2 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 229.0 F 0.0
10:00 216.6 F 0.0
15:00 224.6 F 0.0
20:00 227.3 F 0.0
25:00 211.0 F 0.0
30:00 210.1 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 18:03
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

07/08/2021 02:12 PM
ACCEPTED: X
REJECTED:
Operator: U02