
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001862
13:14 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:14
STERILIZING 4:21
136.0 C 220kPa 4:21
136.4 C 225kPa 4:51
136.4 C 226kPa 5:21
136.4 C 225kPa 5:51
136.1 C 224kPa 6:21
136.4 C 226kPa 6:51
136.2 C 223kPa 7:21
136.4 C 226kPa 7:51
136.4 C 226kPa 7:51
VENTING 7:52
AIR DRYING 8:45
CYCLE COMPLETE 8:59

Digital Signature #
E88292B49C6168FF
