

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3984

Name: RITTER 1  
Sterilizer ID: V2178198  
11/06/2023 12:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.7 F 0.0  
2:00 160.4 F 2.4  
4:00 172.3 F 5.9  
6:00 190.6 F 9.7  
8:00 222.8 F 8.0  
10:00 225.5 F 7.8  
12:00 236.5 F 12.4  
14:00 235.9 F 9.1  
16:00 240.1 F 10.2  
18:00 246.1 F 13.2  
20:00 257.5 F 19.4  
22:00 268.1 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.4  
1:00 271.9 F 28.9  
2:00 272.1 F 28.9  
3:00 272.3 F 29.3  
4:00 271.4 F 28.7

Min 270.9 F 28.3  
Max 273.2 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 193.2 F 0.0  
10:00 200.8 F 0.0  
15:00 198.3 F 0.0  
20:00 193.4 F 0.0  
25:00 189.5 F 0.0  
30:00 186.7 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 23:14  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:02:14  
CYCLE COMPLETE

11/06/2023 01:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_