
Midmark M11 - v1.0.5
Total Cycles: 1350

Name: RITTER2
Sterilizer ID: V2313228
07/26/2022 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.5 F 0.0
2:00 184.7 F 0.4
4:00 196.7 F 4.9
6:00 212.8 F 10.4
8:00 227.3 F 6.8
10:00 241.3 F 12.9
12:00 240.3 F 11.3
14:00 239.4 F 10.4
16:00 253.2 F 17.3
18:00 264.6 F 24.3
20:00 271.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.8 F 30.1
2:00 272.8 F 29.5
3:00 271.6 F 29.2
4:00 272.4 F 30.1

Min 271.5 F 29.1
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 230.7 F 0.0
10:00 226.1 F 0.0
15:00 220.7 F 0.0
20:00 219.3 F 0.0
25:00 216.6 F 0.0
30:00 213.6 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

07/26/2022 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____