
Midmark M11 - v1.0.0
Total Cycles: 3955

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2023 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.3 F 0.0
2:00 179.8 F 2.0
4:00 185.5 F 5.4
6:00 191.2 F 9.7
8:00 224.0 F 7.1
10:00 222.1 F 8.4
12:00 235.1 F 13.2
14:00 237.1 F 10.1
16:00 237.4 F 8.8
18:00 247.3 F 14.2
20:00 258.7 F 20.5
22:00 269.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.3
2:00 271.3 F 28.6
3:00 272.5 F 29.4
4:00 271.8 F 28.9

Min 271.2 F 28.6
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 219.6 F 0.0
10:00 218.9 F 0.0
15:00 222.1 F 0.0
20:00 232.4 F 0.0
25:00 230.2 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 23:04
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

10/30/2023 10:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____