

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000393  
11:12 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 3:02  
STERILIZING 4:22  
136.4 C 233kPa 4:22  
136.4 C 233kPa 4:52  
136.2 C 232kPa 5:22  
136.2 C 231kPa 5:52  
136.1 C 231kPa 6:22  
136.4 C 231kPa 6:52  
136.2 C 231kPa 7:22  
136.4 C 231kPa 7:52  
136.4 C 231kPa 7:52  
VENTING 7:52  
AIR DRYING 8:48  
CYCLE COMPLETE 8:49

-----  
Digital Signature #  
99CA017F5F863302  
-----