
Midmark M11 - v1.0.0
Total Cycles: 4117

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2023 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 175.6 F 3.0
4:00 202.5 F 7.2
6:00 213.3 F 12.8
8:00 225.8 F 6.9
10:00 237.3 F 11.8
12:00 235.9 F 8.8
14:00 240.5 F 10.1
16:00 247.4 F 13.6
18:00 259.6 F 20.2
20:00 270.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.2 F 28.1
2:00 272.4 F 28.6
3:00 271.1 F 27.9
4:00 272.2 F 28.3

Min 271.1 F 27.8
Max 272.6 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.9 F 0.0
5:00 238.8 F 0.0
10:00 232.1 F 0.0
15:00 235.8 F 0.0
20:00 231.1 F 0.0
25:00 218.2 F 0.0
30:00 212.6 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

12/14/2023 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____