
Midmark M11 - v1.0.5
Total Cycles: 2508

Name: RITTER2
Sterilizer ID: V2313228
06/13/2023 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 162.6 F 3.2
4:00 175.5 F 7.1
6:00 199.7 F 12.5
8:00 219.1 F 7.2
10:00 232.8 F 12.4
12:00 234.8 F 9.7
14:00 236.2 F 9.0
16:00 248.1 F 15.1
18:00 260.5 F 22.0
20:00 271.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.4 F 29.8
2:00 271.4 F 29.3
3:00 272.5 F 29.8
4:00 271.5 F 29.1

Min 271.4 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 212.3 F 0.0
10:00 232.6 F 0.0
15:00 229.8 F 0.0
20:00 226.4 F 0.0
25:00 223.8 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

06/13/2023 12:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____