
Midmark M11 - v1.0.0
Total Cycles: 5080

Name: RITTER 1
Sterilizer ID: V2178198
11/12/2024 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 154.2 F 0.1
4:00 162.3 F 0.5
6:00 188.9 F 1.4
8:00 206.1 F 5.0
10:00 218.5 F 8.5
12:00 229.5 F 12.7
14:00 231.3 F 7.9
16:00 242.1 F 12.8
18:00 240.0 F 10.4
20:00 236.3 F 8.2
22:00 250.3 F 15.0
24:00 261.4 F 21.3
26:00 271.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.7 F 28.5
2:00 271.5 F 27.8
3:00 272.6 F 28.7
4:00 271.9 F 28.1

Min 271.4 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 213.3 F 0.0
10:00 197.6 F 0.0
15:00 195.7 F 0.0
20:00 196.8 F 0.0
25:00 224.0 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:05:05
CYCLE COMPLETE

11/12/2024 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____