
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000123
12:56 04/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:35
STERILIZING 4:45
136.0 C 220kPa 4:45
136.3 C 225kPa 5:15
136.3 C 223kPa 5:45
136.4 C 225kPa 6:15
136.2 C 222kPa 6:45
136.2 C 224kPa 7:15
136.3 C 224kPa 7:45
136.4 C 225kPa 8:15
136.4 C 225kPa 8:15
VENTING 8:16
AIR DRYING 9:07
CYCLE COMPLETE 24:08

Digital Signature #
CB9681EB4E160874
