

-----  
Midmark M11 - v1.0.5  
Total Cycles: 260

Name: RITTER2  
Sterilizer ID: V2313228  
08/04/2021 06:38 AM

BEGIN PACKS CYCLE  
Temp: 250 Degrees F  
Time: 30 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.1 F 0.0  
2:00 82.1 F 0.0  
4:00 84.6 F 0.1  
6:00 96.7 F 0.3  
8:00 131.5 F 2.5  
10:00 160.6 F 5.6  
12:00 182.8 F 1.2  
14:00 200.7 F 3.3  
16:00 218.3 F 6.5  
18:00 223.7 F 4.0  
20:00 227.0 F 6.2  
22:00 232.7 F 7.3  
24:00 244.1 F 12.3  
26:00 253.2 F 17.1

STERILIZING  
mm:ss Degrees PSI  
0:00 252.1 F 17.0  
1:00 253.4 F 17.2  
2:00 252.3 F 17.1  
3:00 252.7 F 17.3  
4:00 253.2 F 17.1  
5:00 252.1 F 17.0  
6:00 252.1 F 17.3  
7:00 252.1 F 17.3  
8:00 252.7 F 17.3  
9:00 253.0 F 17.3  
10:00 252.3 F 17.3  
11:00 252.3 F 17.3  
12:00 252.2 F 17.3  
13:00 251.9 F 17.1  
14:00 252.0 F 17.0  
15:00 252.6 F 17.1  
16:00 253.3 F 17.3  
17:00 252.3 F 17.3  
18:00 252.2 F 17.1  
19:00 252.2 F 17.1  
20:00 252.5 F 17.2  
21:00 252.6 F 17.3  
22:00 252.1 F 17.0  
23:00 251.7 F 17.2  
24:00 253.2 F 17.3  
25:00 251.9 F 17.1  
26:00 252.3 F 17.1  
27:00 253.1 F 17.3  
28:00 251.5 F 17.2  
29:00 251.7 F 17.2  
30:00 252.1 F 17.3

Min 251.1 F 17.0  
Max 253.6 F 17.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
5:00 207.1 F 0.0  
10:00 213.7 F 0.0  
15:00 225.4 F 0.0  
20:00 231.9 F 0.0  
25:00 202.0 F 0.0

30:00 157.5 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 26:52  
STERILIZING: 30:00  
VENTING: 2:55  
DRYING: 30:00  
TOTAL CYCLE: 01:30:35  
CYCLE COMPLETE

08/04/2021 08:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01