

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2990

Name: RITTER2  
Sterilizer ID: V2313228  
10/10/2023 02:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.7 F 0.0  
2:00 167.9 F 3.1  
4:00 184.9 F 7.6  
6:00 207.3 F 12.9  
8:00 222.8 F 7.2  
10:00 234.1 F 12.7  
12:00 237.1 F 10.4  
14:00 235.8 F 9.0  
16:00 250.7 F 16.2  
18:00 262.9 F 23.3  
20:00 271.9 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.7 F 29.7  
2:00 271.5 F 28.9  
3:00 272.4 F 29.9  
4:00 272.4 F 29.5

Min 271.3 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
5:00 235.1 F 0.0  
10:00 216.6 F 0.0  
15:00 224.5 F 0.0  
20:00 201.5 F 0.0  
25:00 206.7 F 0.0  
30:00 210.6 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:57  
CYCLE COMPLETE

10/10/2023 03:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_