
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004896
11:17 04/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:40
STERILIZING 3:38
136.5 C 222kPa 3:38
136.3 C 221kPa 4:08
136.4 C 223kPa 4:38
136.4 C 221kPa 5:08
136.4 C 223kPa 5:38
136.4 C 222kPa 6:08
136.3 C 224kPa 6:38
136.0 C 221kPa 7:08
136.0 C 221kPa 7:08
VENTING 7:08
AIR DRYING 7:58
CYCLE COMPLETE 10:09

Digital Signature #
1D7581EDA153D653
