
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001487
11:30 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:36
STERILIZING 3:44
136.4 C 224kPa 3:44
136.4 C 226kPa 4:14
136.4 C 226kPa 4:44
136.2 C 224kPa 5:14
136.4 C 226kPa 5:44
136.4 C 226kPa 6:14
136.3 C 225kPa 6:44
136.3 C 225kPa 7:14
136.3 C 225kPa 7:14
VENTING 7:14
AIR DRYING 8:06
CYCLE COMPLETE 8:10

Digital Signature #
FF497D993616BF70
