
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000721
7:22 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:03
STERILIZING 3:14
136.4 C 232kPa 3:14
136.4 C 232kPa 3:44
136.1 C 230kPa 4:14
136.3 C 231kPa 4:44
136.4 C 232kPa 5:14
136.3 C 231kPa 5:44
136.2 C 230kPa 6:14
136.2 C 228kPa 6:44
136.2 C 228kPa 6:44
VENTING 6:44
AIR DRYING 7:36
CYCLE COMPLETE 7:37

Digital Signature #
0DC9072427B69966
