

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001350  
10:22 16/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:32  
PRESSURIZING 3:34  
STERILIZING 4:54  
136.0 C 220kPa 4:54  
136.4 C 223kPa 5:24  
136.3 C 224kPa 5:54  
136.3 C 223kPa 6:24  
136.4 C 226kPa 6:54  
136.3 C 222kPa 7:24  
136.3 C 225kPa 7:54  
136.4 C 226kPa 8:24  
136.4 C 226kPa 8:24  
VENTING 8:24  
AIR DRYING 9:24  
CYCLE COMPLETE 24:24

-----  
Digital Signature #  
38FF17DDC84FC332  
-----