
Midmark M11 - v1.0.0
Total Cycles: 2639

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2022 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.9 F 0.0
2:00 154.4 F 2.2
4:00 177.8 F 5.6
6:00 202.5 F 10.3
8:00 215.4 F 5.3
10:00 229.3 F 9.6
12:00 233.2 F 8.8
14:00 242.0 F 12.4
16:00 240.8 F 10.8
18:00 254.4 F 17.7
20:00 266.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.1 F 28.6
2:00 272.6 F 29.1
3:00 271.7 F 28.3
4:00 272.7 F 29.1

Min 271.3 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 230.1 F 0.0
10:00 216.6 F 0.0
15:00 210.9 F 0.0
20:00 207.1 F 0.0
25:00 209.7 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

12/05/2022 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____