
Midmark M11 - v1.0.0
Total Cycles: 858

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2021 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 167.5 F 2.3
4:00 174.2 F 5.3
6:00 201.4 F 9.4
8:00 214.5 F 6.2
10:00 226.3 F 8.1
12:00 238.2 F 13.3
14:00 238.7 F 10.5
16:00 237.5 F 8.4
18:00 250.1 F 15.1
20:00 261.6 F 21.7
22:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.8
2:00 271.6 F 28.1
3:00 272.2 F 28.9
4:00 272.2 F 28.5

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 212.0 F 0.0
10:00 223.8 F 0.0
15:00 232.7 F 0.0
20:00 231.8 F 0.0
25:00 228.1 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

06/01/2021 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____