
Midmark M11 - v1.0.0
Total Cycles: 2280

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2022 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.3 F 0.0
2:00 205.3 F 1.7
4:00 210.5 F 6.2
6:00 216.3 F 11.7
8:00 226.4 F 6.6
10:00 238.1 F 11.9
12:00 236.7 F 9.2
14:00 240.3 F 9.4
16:00 248.3 F 14.2
18:00 260.2 F 20.8
20:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.3 F 28.9
2:00 271.8 F 28.2
3:00 272.4 F 29.0
4:00 272.0 F 28.4

Min 271.5 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 230.8 F 0.0
10:00 232.0 F 0.0
15:00 227.7 F 0.0
20:00 227.1 F 0.0
25:00 228.8 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

09/02/2022 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____