

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076

Unit # : 000

WATER QUALITY

2.0uS / 1.2ppm

CYCLE NUMBER 000025

8:28 14/01/2020

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 4:19

PRESSURIZING 10:10

STERILIZING 11:19

123.5 C 121kPa 11:19

123.4 C 120kPa 11:49

123.4 C 120kPa 12:19

123.4 C 120kPa 12:49

123.4 C 120kPa 13:20

123.5 C 120kPa 13:50

123.3 C 120kPa 14:20

123.4 C 120kPa 14:50

123.5 C 120kPa 15:20

123.4 C 120kPa 15:50

123.4 C 120kPa 16:20

123.4 C 120kPa 16:50

123.3 C 120kPa 17:20

123.2 C 120kPa 17:50

123.3 C 120kPa 18:20

123.4 C 120kPa 18:50

123.5 C 120kPa 19:20

123.5 C 120kPa 19:50

123.4 C 119kPa 20:20

123.4 C 120kPa 20:50

123.5 C 120kPa 21:20

123.4 C 120kPa 21:50

123.4 C 120kPa 22:20

123.3 C 120kPa 22:50

123.3 C 119kPa 23:20

123.4 C 120kPa 23:50

123.5 C 120kPa 24:20

123.3 C 120kPa 24:50

123.4 C 120kPa 25:20

123.4 C 120kPa 25:50

123.5 C 120kPa 26:19

VENTING 26:20

AIR DRYING 27:22

CYCLE COMPLETE 57:22

-----

Digital Signature #

355B7855D1038835

-----