
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001159
11:22 08/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:29
STERILIZING 3:29
136.0 C 220kPa 3:29
136.4 C 225kPa 3:59
136.3 C 225kPa 4:29
136.4 C 225kPa 4:59
136.4 C 225kPa 5:29
136.3 C 225kPa 5:59
136.4 C 225kPa 6:29
136.4 C 224kPa 6:59
136.4 C 224kPa 6:59
VENTING 6:59
AIR DRYING 7:52
CYCLE COMPLETE 22:52

Digital Signature #
27AED8CA5E1704CC
