
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002990
15:27 19/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:28
STERILIZING 3:32
136.4 C 220kPa 3:32
136.1 C 222kPa 4:02
136.4 C 221kPa 4:32
136.3 C 224kPa 5:02
136.4 C 225kPa 5:32
136.3 C 224kPa 6:02
136.2 C 219kPa 6:32
136.2 C 222kPa 7:02
136.2 C 222kPa 7:02
VENTING 7:03
AIR DRYING 7:56
CYCLE COMPLETE 8:06

Digital Signature #
4F7AAB2F3FFE31C3
