

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1431

Name: RITTER2  
Sterilizer ID: V2313228  
08/30/2022 09:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
2:00 187.5 F 1.3  
4:00 204.7 F 5.4  
6:00 221.2 F 10.8  
8:00 228.0 F 5.8  
10:00 236.3 F 11.1  
12:00 236.2 F 8.7  
14:00 241.4 F 10.7  
16:00 247.0 F 14.3  
18:00 259.7 F 21.3  
20:00 271.0 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.6 F 29.5  
2:00 272.0 F 29.5  
3:00 272.4 F 29.4  
4:00 272.1 F 29.5

Min 271.6 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 197.5 F 0.0  
10:00 213.3 F 0.0  
15:00 231.1 F 0.0  
20:00 220.0 F 0.0  
25:00 224.0 F 0.0  
30:00 214.4 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 20:41  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:36  
CYCLE COMPLETE

08/30/2022 10:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_