
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000795
11:15 30/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:12
STERILIZING 3:07
136.0 C 228kPa 3:07
136.4 C 231kPa 3:37
136.3 C 232kPa 4:07
136.4 C 233kPa 4:37
136.4 C 232kPa 5:07
136.4 C 233kPa 5:37
136.1 C 231kPa 6:07
136.3 C 232kPa 6:37
136.3 C 232kPa 6:37
VENTING 6:37
AIR DRYING 7:30
CYCLE COMPLETE 7:35

Digital Signature #
0DCF66DA88F6F788
