
Midmark M11 - v1.0.0
Total Cycles: 1662

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2022 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.5 F 0.0
2:00 184.4 F 3.7
4:00 200.7 F 8.1
6:00 217.1 F 10.0
8:00 227.6 F 8.6
10:00 238.0 F 11.0
12:00 242.0 F 12.0
14:00 240.5 F 10.6
16:00 254.2 F 17.6
18:00 265.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.2
2:00 272.0 F 28.6
3:00 271.5 F 28.4
4:00 272.6 F 29.2

Min 271.3 F 28.2
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 224.9 F 0.0
10:00 230.3 F 0.0
15:00 225.9 F 0.0
20:00 226.1 F 0.0
25:00 231.1 F 0.0
30:00 230.6 F 0.0

Drying Complete

FILLING: 1:49
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

02/23/2022 03:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____