
Midmark M11 - v1.0.5
Total Cycles: 1585

Name: RITTER2
Sterilizer ID: V2313228
10/12/2022 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.8 F 0.0
2:00 181.6 F 3.7
4:00 196.5 F 7.8
6:00 220.1 F 13.1
8:00 226.9 F 7.9
10:00 238.2 F 13.3
12:00 237.7 F 11.1
14:00 237.7 F 9.9
16:00 251.8 F 16.8
18:00 263.7 F 23.9
20:00 271.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 29.8
2:00 271.6 F 29.0
3:00 271.9 F 30.0
4:00 272.1 F 29.6

Min 271.2 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 221.4 F 0.0
10:00 226.7 F 0.0
15:00 226.8 F 0.0
20:00 224.2 F 0.0
25:00 225.7 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

10/12/2022 01:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____