
Midmark M11 - v1.0.5
Total Cycles: 943

Name: RITTER2
Sterilizer ID: V2313228
04/01/2022 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 195.0 F 3.0
4:00 208.1 F 8.7
6:00 227.1 F 9.5
8:00 234.5 F 11.5
10:00 236.6 F 10.0
12:00 236.7 F 9.4
14:00 251.7 F 16.9
16:00 264.1 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.6 F 29.8
2:00 271.6 F 29.5
3:00 272.6 F 29.8
4:00 271.5 F 29.2

Min 271.4 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 215.7 F 0.0
10:00 228.0 F 0.0
15:00 234.1 F 0.0
20:00 230.8 F 0.0
25:00 225.4 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:50
HEATING: 17:55
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

04/01/2022 10:55 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____