
Midmark M11 - v1.0.0
Total Cycles: 2289

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2022 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.1 F 0.0
2:00 138.9 F 0.2
4:00 152.4 F 1.8
6:00 174.6 F 4.1
8:00 196.8 F 8.9
10:00 218.2 F 10.1
12:00 229.5 F 7.7
14:00 239.8 F 12.7
16:00 239.1 F 9.9
18:00 237.2 F 8.4
20:00 250.0 F 14.9
22:00 262.0 F 21.7
24:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.4 F 28.5
2:00 271.2 F 27.8
3:00 272.4 F 28.7
4:00 271.9 F 28.0

Min 271.2 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 212.7 F 0.0
10:00 227.3 F 0.0
15:00 230.8 F 0.0
20:00 245.3 F 0.0
25:00 232.8 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:42
CYCLE COMPLETE

09/06/2022 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____