
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000748
7:14 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:16
PRESSURIZING 1:23
STERILIZING 2:19
136.0 C 220kPa 2:19
136.2 C 231kPa 2:49
136.3 C 230kPa 3:19
136.4 C 232kPa 3:49
136.2 C 230kPa 4:19
136.3 C 231kPa 4:49
136.4 C 232kPa 5:19
136.3 C 230kPa 5:49
136.3 C 230kPa 5:49
VENTING 5:50
AIR DRYING 6:42
CYCLE COMPLETE 6:46

Digital Signature #
5BBF8C925A338ABD
