
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001318
10:49 26/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:43
STERILIZING 4:57
136.0 C 220kPa 4:57
136.2 C 224kPa 5:27
136.3 C 224kPa 5:57
136.4 C 225kPa 6:27
136.4 C 223kPa 6:57
136.4 C 224kPa 7:27
136.3 C 224kPa 7:57
136.4 C 224kPa 8:27
136.4 C 224kPa 8:27
VENTING 8:27
AIR DRYING 9:25
CYCLE COMPLETE 24:25

Digital Signature #
A3D569AE91B2938F
