
Midmark M11 - v1.0.0
Total Cycles: 5106

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2024 06:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.0 F 0.0
2:00 169.6 F 0.4
4:00 197.3 F 2.9
6:00 222.4 F 8.2
8:00 231.0 F 12.5
10:00 235.7 F 9.4
12:00 237.1 F 9.1
14:00 246.8 F 13.2
16:00 245.3 F 12.6
18:00 258.6 F 19.7
20:00 270.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.7 F 28.6
2:00 271.6 F 27.9
3:00 272.8 F 28.7
4:00 271.7 F 27.9

Min 271.5 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 199.9 F 0.0
10:00 206.9 F 0.0
15:00 231.4 F 0.0
20:00 246.0 F 0.0
25:00 228.5 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

11/18/2024 07:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____