
Midmark M11 - v1.0.5
Total Cycles: 3896

Name: RITTER2
Sterilizer ID: V2313228
08/27/2024 02:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.1 F 0.0
2:00 108.0 F 0.7
4:00 139.1 F 3.6
6:00 167.0 F 8.2
8:00 195.5 F 12.7
10:00 216.3 F 6.9
12:00 230.5 F 12.0
14:00 233.4 F 8.9
16:00 239.6 F 10.0
18:00 246.9 F 13.9
20:00 259.5 F 20.9
22:00 271.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 271.8 F 28.8
2:00 271.8 F 28.9
3:00 271.8 F 28.9
4:00 271.9 F 28.7

Min 271.7 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 230.4 F 0.0
5:00 216.6 F 0.0
10:00 228.2 F 0.0
Dry Cycle Aborted

FILLING: 0:45
HEATING: 22:35
STERILIZING: 4:00
VENTING: 2:48
DRYING: 12:45
TOTAL CYCLE: 00:42:53

08/27/2024 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____