
Midmark M11 - v1.0.5
Total Cycles: 1987

Name: RITTER2
Sterilizer ID: V2313228
01/31/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.3 F 0.0
2:00 193.3 F 4.4
4:00 206.7 F 10.1
6:00 220.7 F 6.3
8:00 233.7 F 11.6
10:00 236.1 F 9.6
12:00 237.3 F 9.1
14:00 250.5 F 15.8
16:00 262.9 F 23.0
18:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.4
2:00 271.4 F 28.8
3:00 272.6 F 29.5
4:00 271.7 F 28.8

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 222.6 F 0.0
10:00 232.6 F 0.0
15:00 234.1 F 0.0
20:00 228.1 F 0.0
25:00 224.7 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

01/31/2023 02:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____