

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5341

Name: RITTER 1  
Sterilizer ID: V2178198  
02/19/2025 09:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.5 F 0.0  
2:00 77.8 F 0.3  
4:00 89.6 F 1.9  
6:00 125.1 F 4.2  
8:00 154.2 F 6.8  
10:00 177.8 F 10.0  
12:00 189.8 F 13.5  
14:00 204.0 F 3.9  
16:00 216.1 F 7.3  
18:00 227.6 F 11.0  
20:00 226.8 F 6.4  
22:00 238.2 F 10.8  
24:00 234.6 F 7.6  
26:00 247.8 F 14.1  
28:00 259.0 F 20.1  
30:00 269.1 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.5 F 28.6  
2:00 272.1 F 28.7  
3:00 272.0 F 28.3  
4:00 272.6 F 29.0

Min 271.5 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.9 F 0.0  
5:00 177.7 F 0.0  
10:00 190.9 F 0.0  
15:00 197.9 F 0.0  
20:00 204.7 F 0.0  
25:00 210.9 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 31:04  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:09:21  
CYCLE COMPLETE

02/19/2025 10:37 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_